



सीएसआईआर-केंद्रीय खाद्य प्रौद्योगिक अनुसंधान संस्थान
CSIR- CENTRAL FOOD TECHNOLOGICAL RESEARCH INSTITUTE
मैसूरु / MYSURU-570 020, भारत / INDIA

(Constituent Laboratory of CSIR, New Delhi (Ministry of Science & Technology)
An ISO 9001:2008, ISO 14001:2004 & ISO 17025:2005, NABL Accredited Laboratory

Corrigendum: Tender for High Performance Liquid Chromatography (HPLC)

Corrigendum Title: Revised Technical Specifications based on PBC

Tender Ref: CFTRI/52361/25-26 Date: 07-04-2025

Tender ID: 2025_CSIR_232222_1

The revised final specifications based on the deliberations in Pre Bid Conference held on 22-04-2025 @ 11. 00A.M is uploaded herewith.

All the prospective bidders are requested to take cognizance of the revised specifications and submit their bids accordingly on or before 03.00 p.m. on

13-05-2025.

All other tender terms and conditions of tender remain unaltered.

Controller of Stores &Purchase

CSIR-CFTRI, Mysore

Dt. 30-04-2025

Revised Technical Specification based on PBC

SPECIFICATION FOR HIGH PERFORMANCE LIQUID CHROMATOGRAPH [HPLC]:

- The HPLC system should have the following individual stackable self-contained HPLC modules.
- They must be capable of standalone operation and have internal diagnostics to monitor the instrument's performance and record in a logbook.
- Modules must be connected with high-speed transmission technology to enhance the reliability and sensitivity of the HPLC.
- The instrument will be used for method development for nutraceuticals; hence the system should have the provision to accommodate minimum four columns along with automation control. The instrument should select the column automatically based on the program.

- 1. Quaternary Gradient Pump**
- 2. Auto-Sample Injector with Accessories**
- 3. Column Oven & Columns**
- 4. Photodiode Array Detector**
- 5. Fluorescence Detector**
- 6. Data Management System, Printer & Software**

1. Quaternary Gradient Pump

- It should be Quaternary Gradient pump.
- The flow rate should be set between 0.001 to 10 ml/min.
- Flow rate accuracy should be $\pm 1\%$ or ± 0.5 $\mu\text{l}/\text{min}$ of set value whichever is larger
- Flow rate precision should be less than $\leq 0.07\%$ RSD
- Maximum Pressure: 400 bar or better
- The standard gradient mixer should have a pump delay volume of 650 μl .
- The composition precision should be less than $\pm 0.2\%$
- Maintenance kit, reservoir tray (with 4 solvent bottles complete with fittings) & automatic rinsing kit must be supplied
- It should have functions for maintenance and validation which are accessible by a dedicated operation button.
- Membrane degassing unit should have minimum three flow lines with maximum operating flow rate should be up to 10 ml/min per flow line
- It must have a leak sensor as a safety feature and error status shall be transferred to the operating software
- It should include a drain-pan and should be equipped with self-cleaning capability that extends vacuum pump life by drawing in air when the pump is running.

2 Auto-Sample Injector with Accessories

- The autosampler design should be a flow through needle design with variable injection volume with zero sample loss during injection
- Volume: 0.1 μl to 100 μl
- Loop injection using fixed vol or variable vol loop
- Injection time: Minimum < 20 sec/sample or better
- It should be able to use four different solvents
- Temperature setting range : 4 to 40°C
- The carry over must be below 0.004 %
- The injection precision should be less than 0.25% of RSD
- Number of samples to be processed automatically, random access upto for 1ml, 1.5ml, 96, 384 wells microtitre plates or suitable higher capacity with sample preparation for auto-dilution.
- Dedicated operational key functions for maintenance and validation

3. Column Oven

- It should be forced-air-circulation/ peltier type for uniform temperature distribution with a quick feedback mechanism to maintain constant temperature
- The temperature range : 4 - 85°C \pm 0.1°C precision
- The oven should have a solvent leak sensor
- Dedicated operational key functions for maintenance and validation
- It should be able to handle up to 3 x 30 cm columns or better Columns: The following column should be supplied along with system :
- C18 column (5 μ m x 4.6mm x 250mm)
- C18 column (3 μ m x 4.6mm x 150mm)
- HILIC column (5 μ m x 4.6mm x 250mm)

4. Photodiode Array Detector

- The wavelength range should be minimum 190 nm - 800 nm
- The photo-diode array detector should have 1024 elements and an element resolution of 1.4 nm/element or better must be available.
- Standard Cell: Optical path length:10mm, Capacity:8-13 μ l , Pressure:8MPa
- Operating temperature setting range 4 to 35°C
- Wavelength accuracy should be \pm 1 nm
- The Drift must be $\leq 0.9 \times 10^{-3}$ AU/Hour
- The Noise Level $\leq 7 \times 10^{-6}$ AU
- Linearity should be equal or more than 2.0AU
- Sampling rate should be more than 100Hz
- Peak purity function should be available.
- It should have automatic wavelength accuracy check at 4 wavelengths (UV & Vis) & wavelength correction

5. Fluorescence Detector

- Light source should be Xenon lamp
- Wavelength range should be 200 nm to 650 nm or better
- Simultaneous monitoring of four wavelength should be possible
- Spectral bandwidth should be 20 nm or better
- Flow cell volume should be 12 μ l or lower
- Wavelength accuracy should be : \pm 3nm or better
- Wavelength repeatability: \pm 0.2nm or better
- Signal to noise ratio equal to or more than 500 RL(Raman Line)

6. Data Management System & software

- Suitable system with minimum Intel Core i7 (14700), 1 TB SDD, 64GB (DDR5) RAM, LAN card, 24" LED (FHD) Monitor, Brand Preferred: Dell /HP
- Duplex Laserjet printer with scanner

- It should cover full digital instrument control, qualitative and quantitative processing, report creation, self-diagnosis and auto-tuning.
- Full 2D and 3D spectral analysis for Diode-Array Detector should be available.
- The self-diagnosis feature of the software should enable diagnosis of all detectors and all connected units.
- The software must have 'column management' system for recording the column details and column usage, perform should perform customised continuous analyses according to the conditions specified for each sample.
- The software should be capable to perform overlapping injections, analyze safety feature where in when the mobile phase level goes down below the set threshold value, the software should trigger auto shutdown of pump, autopurge
- The software should be able to monitor and quantitate minimum 8 different wavelengths simultaneously.
- The data can be converted to other formats. Spread Sheet software and word-processing software can be readily employed to provide data in tables or graphs through industry standard protocols.
- System suitability, System security as well as System check functions must be provided which comply with Good Laboratory Practice (GLP) and Regulatory Conformity.

Service, Warranty and Training

- Minimum one-year comprehensive warranty/guarantee for the entire system with accessories, from the date of satisfactory installation and commissioning
- The warranty shall commence only upon successful completion of the Acceptance Test or commissioning.
- On-site installation, commissioning and training shall be conducted by a qualified factory-trained engineer.
- Vendor to provide service guarantee: should the system require service during the warranty period, vendor must guarantee or replacement of instrument for free.
- Vendor to have logistic support over atleast 10years or more.
- Vendor to provide a copy of Site-Preparation checklist.
- Vendor to provide both on-site and operator training for users on the system start-up, usage, maintenance, quality control, trouble shooting, etc. including comprehensive classroom training.
- Vendor must demonstrate that it has a proven appropriate set-up and capability to provide after-sales service efficiently and effectively. The supplier should have in his facility a similar system to that proposed in this tender for training purpose.

Additional requirements

- Printed literature of the manufacturer is to be provided in support of the technical specifications.
- A point wise compliance statement as per the specifications must be submitted along with the offer. & Specify the site requirement and must include pre-installation site inspection.
- Must provide list of Govt institute and installations in last five years with the contact details of user scientists.
- Post warranty AMC charges for 5 years.